

TESTING KITS FOR DRUGS

Procedure Card for IDT2400T

OPIATE IDENTIFICATION

Substance & Trace Test



Read Before Use As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not prove probe once a capsule is broken. In the unlikely event of a splil, rinse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.



Sample suspected substance or surface.
Substance & Trace Material
Swipe the paper of the

Swipe the paper of the Smart-TipTM probe across the suspect substance or surface

Lock Smart-Tip™ probe into test unit. Press firmly.



Do not continue unless probe is locked below the Flex Pins and cannot be removed from the test unit.

dentification



Gently shake test unit for 2-3 seconds.



Results



The appearance of a purple color <u>within one minute</u>, in either the front or back window, indicates the presence of an opiate such as **Heroin**, **Morphine** or **Codeine**.




If no purple color develops, an opiate, such as Heroin, Morphine or Codeine, is <u>not</u> present.



Proceed to Dilution (steps 6-9), regardless of whether or not a color appeared.



ution






Gently shake test unit for 2-3 seconds.



Wait 5 minutes.
The test window will warm up.
This is normal.

Dilution





Gently shake test unit for 2-3 seconds.



The test window will warm up.
This is normal.

甸

Disposal

After Dilution, the test kit is safe to dispose of in the trash.



Manufactured by IDenta Ltd

www.identa-corp.com Patented Technology Made in Israel