














Read Before Use



As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not puncture any part of unit. Do not remove probe once a capsule is broken. In the unlikely event of a spill, rinse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.

Test Preparation	1	Remove the Smart-Tip™ probe from test unit.	2 Sample suspected substance.	
			Powder Touch Smart-Tip™ probe to sample. 	Tablet Firmly press Smart-Tip™ probe on the tablet while scrubbing in a circular motion until the tip is completely covered with substance. 

Test Preparation	3	Lock Smart-Tip™ probe into test unit. Press firmly. 	STOP Do not continue unless probe is locked below the Flex Pins and cannot be removed from the test unit.	Identification	4	With thumb and forefinger, squeeze chamber 1 until capsule breaks. 	5	Gently shake test unit for 2-3 seconds. 
	Flex Pins							

Results		The appearance of a violet color strongly indicates the presence of Ecstasy/MDMA or its analogues. * The color may take up to 3 minutes to develop.
		The appearance of a yellow color indicates the presence of Bath Salts such as MDPV, Methyone or Butylone.
		If none of the above colors develop, the above listed drugs are not present.
		Proceed to Dilution (steps 6-9), regardless of whether or not a color appeared.
		

Dilution	6	With thumb and forefinger, squeeze chamber 2 until capsule breaks. 	7	Gently shake test unit for 2-3 seconds. 	Wait 5 minutes. The test window will warm up. This is normal.
----------	---	--	---	---	---

Dilution	8	With thumb and forefinger, squeeze chamber 3 until capsule breaks. 	9	Gently shake test unit for 2-3 seconds. 	The test window will warm up. This is normal.
----------	---	--	---	---	---

Disposal After Dilution, the test kit is safe to dispose of in the trash.